



May 26, 2020

Travel Compliance Letter for Aria Health Therapy Devices

To whom it may concern,

Pneumatic compression devices (PCDs) are medical devices prescribed for use on or by the order of a physician and are cleared for sale by the U.S Food and Drug Administration, a branch of the federal government of the United States of America.

The U.S. Department of Transportation (DOT) Final Rule, "Non-discrimination on the Basis of Disability in Air Travel" (73 FR 27614 which updates Title 14 CFR Part 382), effective May 13, 2009, provides requirements for the accommodation of passengers with a disability. These requirements describe allowances for assistive devices, which include medical devices such as PCDs.

The device does not contain any parts or materials known to cause or interfere with the safe operation of a commercial aircraft.

Aria Health has successfully completed testing for the device described in this letter. **The devices listed below have been tested for electromagnetic compatibility (EMC) in accordance with RTCA/DO-160G, Section 21, Category M, and are considered FAA compliant.** *This information may be engraved on the device or a sticker label applied to the device; either method for identifying FAA compliance is acceptable and in accordance with DOT regulations.*

Aria Health product	Product codes
Aria Free™	R0001, R0002, R0003, R0004

Please contact Aria Health customer service with any questions.

United States: +1 (866) 647 0691 (Toll free)

Kind regards,

Eric Saunders
Director, Operations and Regulatory
Eric.Saunders@aria-therapy.com